

OIML Member State
The Netherlands

Number R76/2006-NL1-16.33
Project number 15200575
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Farasoo Towzin Eng Co. 11th Street , Mamoonieh Industrial City Saveh Iran
Identification of the certified type	An Indicator Type : GAMMA-07, GAMMA-07C
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class or

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
28 April 2016

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).





OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-16.33
Project number 15200575
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Report(s):

- No. NMI-15200575-01 dated 22 April 2016 that includes 52 pages.

Characteristics of the indicator:

Accuracy class	III or IIII
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 10000$ divisions
Load cell excitation voltage	5 V square wave (when no load cell connected)
Minimum input voltage per verification scale interval	0,5 μ V
Minimum load cell resistance	35 Ω
Maximum load cell resistance	1236 Ω
Fraction of the maximum permissible error	0,5
Load cell connection	6-wire (remote sensing)
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	2416 m/mm ²
Temperature range	-10 °C / +40 °C
Power supply voltage	160 – 240 V AC 50/60 Hz
Software identification	Version number: 2.6